

File Revision Date:

November 26, 2024

License choice:

CC-BY-SA 4.0 license (open – but credits required – share alike)

Data Set Description:

PI: Hagen Telg
Associate Investigator: John E. Barnes
Instrument: Lidar
Site(s): Mauna Loa Observatory
Measurement Quantities: Aerosol Backscatter (532 nm, 1064 nm), water vapor and Temperature

Contact Information:

Name: Hagen Telg
Address: NOAA/Global Monitoring Laboratory
325 Broadway, Boulder, CO 80303
Phone: (505) 205-5425 cell
Email: hagen.telg@noaa.gov

Name: John E. Barnes
Address: NOAA/Global Monitoring Laboratory
325 Broadway, Boulder, CO 80303
Phone: (720) 684-9143 cell
Email: John.E.Barnes@noaa.gov

Reference Articles:

Determination of Aerosol Height Distributions by Lidar, F.G. Fernald, B.M. Herman, and J.A. Reagan, Journal of Applied Meteorology, Vol. 11, April, 1972.

Methodology for error analysis and simulation of lidar aerosol measurements, P.B. Russell, T.J. Swissler, and M.P. McCormick, Applied Optics, Vol. 18, No. 22, November, 1979.

Correction for nonlinear photon-counting effects in lidar systems, D.P. Donovan, J.A. Whiteway, and A.I. Carswell, Applied Optics, Vol. 32, No. 33, November, 1993.

Barnes, J. E., and D. J. Hofmann, Lidar measurements of stratospheric aerosol over Mauna Loa Observatory, Geophys. Res. Lett., 24, 1923-1926, 1997.

Barnes, J. E., and D. J. Hofmann, Variability in the stratospheric background aerosol over Mauna Loa Observatory, Geophys. Res. Lett., 28, 2895-2898, 2001.

David Hofmann, John Barnes, Michael O'Neill, Michael Trudeau, and Ryan Neely, Increase in background stratospheric aerosol observed with lidar at Mauna Loa Observatory and Boulder, Colorado, Geophys. Res. Lett., 36, doi:10.1029/2009GL039008, 2009.

Stefanie Kremser, et. al, Stratospheric aerosol-observations processes, and impact on climate (2016), Rev. Geophysics, 54,10.1002/2015RG000511.

Christine Bingen, et. al, Stratospheric aerosol data records for the climate change initiative: Development, validation and application to chemistry-climate modelling (2017), Remote Sensing of Environment, 203, 296-321, 10.1016/j.rse.2017.06.002.

Yu, Pengfei, Sean M. Davis, Owen B. Toon, Robert W. Portmann, Charles G. Bardeen, John E. Barnes, Hagen Telg, Christopher Maloney and Karen H. Rosenlof, (2021), Persistent Stratospheric Warming Due to 2019–2020 Australian Wildfire Smoke, Geophysical Research Letters, 48, 7, 10.1029/2021GL092609

Sergey Khaykin, et. al, Global perturbation of stratospheric water and aerosol burden by Hunga eruption, Communications Earth & Environment (2022), doi.org/10.1038/s43247-022-00652-x.

Instrument Description:

The lidar is based on a Spectra Physics GCR-6 Nd:YAG laser (30 Hz, 40 Watt) with frequency doubling and tripling (1064, 532, 355 nm) although the 355 nm has not been used routinely. One 61 cm telescope is dedicated to 532 nm measurements, one 61 cm to the 1064 nm and a 74 cm to Raman nitrogen (607 nm) and water vapor (660 nm). Photomultiplier tubes are used in the photon counting mode for all channels. The system uses a PC 80486 and the data acquisition electronics are MSC II boards made by Tennelec. The lidar control program is written in Turbo Pascal which is needed due to the MCS data acquisition electronics running on the MS/DOS operating system. Measurements are made during the night, usually once a week.

Algorithm Description:

The calculation is a single pass calculation (Fernald et al, 1972) which uses a reference altitude range (no aerosol) between 38 and 40 km. This entire range is forced to an average aerosol backscatter of zero, with the altitude bins weighted by signal error. The atmospheric density comes from the Hilo, Hawaii National Weather Service radiosondes and the MAP 85 model interpolated in time from monthly averages above the radiosonde.

In 2012 the analysis was modified include a better interpolation of atmospheric density between the radiosonde and the MAP85 model. The change was slight but effected the very low aerosol loading during background conditions, especially in the winter. The decrease of winter values reduced the annual cycle published in Barnes and Hofmann, 2001. The database has been reanalyzed with the new interpolation method.

Between 2017 and 2021 the analysis software was rewritten in Python using the Anaconda implementation.

Expected Precision/Accuracy of Instrument:

The error in the backscatter ratio is contained in the data file at each altitude. The typical error is about 5% for the integrated aerosol stratospheric backscatter. This error (Russell et al, 1979) accounts for the signal error both at the altitude being calculated and the reference altitude, and the error resulting from the transmission calculation. The aerosol measurement has been intercompared with the Ruby lidar, also operated by NOAA/Mauna Loa Observatory, and with the NASA/LITE instrument. The temperature measurement has been intercompared with three other lidars during the MLO3 campaign in August, 1995. Further aerosol and water vapor comparisons have been made with the NASA/Goddard mobile NDACC lidar.

Instrument History:

Aerosol backscatter was first measured in April, 1994. Temperature measurements started in June, 1994. Routine water vapor measurements started in October, 2005. Measurements are usually made weekly, more often during intensive campaigns.

Measurements continued until the eruption of Mauna Loa in November, 2022 which covered part of the road and cutoff the electricity. The measurements will hopefully resume in late 2025.